



Osseointegration Monitoring Device

Osseo 100+

English

OPERATION MANUAL

Please read this OPERATION MANUAL carefully
Before use and file for future reference



MADE IN SWEDEN

55185 Rev. 0, February 2020





Fig 1



Fig 2



Fig 3



Fig 4



Fig 5

1. Indications for Use

Osseo 100+ is indicated for measuring the stability of dental implants in the oral cavity or craniofacial region.

2. Intended users


Professional health care users and Professional health care facility environments only. Please read the instruction for use before the first usage.


3. Figures and System components


Fig 1 Osseo100+ Instrument	Included in package
Fig 2 MultiTipeg Driver	Included in package
Fig 3 Example MultiTipeg	Not included, sold separately
Fig 4 Mains adapter and plugs	Included in package
Fig 5 Measurement position	Shows how the instrument tip is held towards the MultiTipeg during a measurement


4. Specifications

- Power input: 5VDC, 1 VA
- Charger input: 100-240 VAC, 5VA
- Instrument weight: 78g
- Charger safety class: EN 60601-1 Class II
- Instrument safety class: EN 60601-1 ME Class II
- EMC: EN 60601-1-2, class B
- The instrument is intended for continuous use
- The instrument contains NiMH batteries.

 Only original parts should be used

 Power supply: Use only the supplied mains adapter and plugs

 No user modification of this equipment is allowed

 Batteries should be collected separately



















5. Operating environment

Ambient temperature: -20° to 65° C (-4°-149° F)
Relative humidity: 45% to 85% Rh, non-condensing

6. Transport and storage

Ambient temperature: -20° a 45° C (-4°-113° F). Relative humidity: 45%-85% Rh.

7. Symbols


	Warning		Keep dry
	Follow instructions for use		Temperature limit
	Magnetic field warning		Manufacturer
	Autoclavable up to 134° C		Manufacturing year
	Delivered Non-sterile		CE mark
	Catalog number		Caution: Federal law restricts this device to sale by or on the order of a physician or dentist.
	Lot/Batch code		Waste from electronic equipment must be handled according to local regulations
	Serial number		Type BF Applied part The instrument and MultiTipeg
	Bluetooth technology		Federal Communications Commission (FCC) approved equipment. FCC ID: 2AEMXY7011A00000, IC 20194-Y7011A00000

8. Characteristics

Osseo 100+ is an instrument for measuring the stability (ISQ) of dental and craniofacial implants. The instrument measures the resonance frequency of a measurement pin "MultiTipeg" and presents it as an ISQ value. The ISQ value, 1-99, reflects the stability of the implant – the higher the value, the more stable the implant.

The instrument measures the ISQ-value with a precision of +/- 1 ISQ unit. When mounted onto an implant, the MultiTipeg resonance frequency can vary up to 2 ISQ units depending on the tightening torque.


By connecting Osseo 100+ to a Surgic Pro2 (NSK) device, the ISQ value and battery status are transferred. For more information, see the Surgic Pro2 manual and "14. Usage" below.

 Warning: Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation.

9. MultiTipeg

The MultiTipeg is made from titanium, and has an integrated grip for the driver on top. Inspect the MultiTipeg for damage before use. Damaged MultiTipegs should not be used due to the risk of erroneous measurements.

There are different MultiTipegs available made to fit different implant systems and types. Please refer to the updated list from the supplier.

 Measurements should only be performed using the correct MultiTipegs. Using the wrong MultiTipeg could cause erroneous measurements or damages to the MultiTipeg or implant.



The instrument emits short magnetic pulses with pulse duration of 1 ms and strength of +/- 20 gauss, 10 mm from the instrument tip. Precautions might be necessary when using the instrument close to cardiac pacemakers or other equipment sensitive to magnetic fields.

10. Technical function

For bringing the measurement pin into vibration, short magnetic pulses are sent from the instrument tip. The magnetic pulses interact with the magnet inside the measurement pin and cause the pin to vibrate. A pickup in the instrument picks up the alternating magnetic field from the vibrating magnet, calculate the frequency and from that, the ISQ value.

11. Implant stability

An implant can have different stabilities in different directions. Make sure to measure from different directions around the top of the MultiPeg.

12. ISQ-value

The stability of the implant is presented as an "ISQ value". The higher the value, the more stable the implant. The ISQ is described in numerous clinical studies. A list of studies can be ordered from the supplier.

13. Batteries & charging

The instrument contains 2 NiMH battery cells that have to be charged before use. A full charge takes approximately 3 hours. From fully charged, the instrument can measure continuously for 60 minutes before it needs to be recharged. The yellow LED is lit when the battery needs recharging. When the battery reaches a critical level, the instrument shuts off automatically. When the batteries are charging, the blue LED is lit. When the batteries are fully charged, the light goes off. The charger should not be plugged in while measuring due to the risk of power line interference making it difficult to measure.

14. Usage

14.1 Instrument on/off

To turn the instrument on, press the operating key. A short beep should be heard and then all display segments are lit up for a short while. Check that all display segments are lit.

The software version is then shown briefly before the instrument starts to measure. If any error code (EX, where "X" is the error number) is shown during start up, please refer to the section "Troubleshooting".

To turn off, press and hold the operating key until the instrument turns off. The instrument will power down automatically after 1 minute of inactivity.

14.2 Measurement

A MultiPeg (fig 3) is mounted onto the implant by using the MultiPeg driver (fig 2). Use hand-tightening with 6-8 Ncm of tightening torque Turn on the instrument and hold the tip close to the top of the pin (fig 5). When a signal is received, a beep sound is heard and then the ISQ-value is shown on the display for a short while before the instrument starts to measure again.

If electromagnetic noise is present, the instrument cannot measure. The electromagnetic noise warning is audible as well as visible on the display. Try to remove the source of the noise. The source could be any electric equipment close to the instrument. When an ISQ value is displayed, it is simultaneously sent via Bluetooth to a Surgic Pro2 (NSK) device if a Bluetooth connection has been made (See 14.3.1).

14.3 ISQ Bluetooth transfer

14.3.1 Bluetooth pairing

To establish the Bluetooth data transfer, the instrument has to be paired with a Surgic Pro2 device. The pairing only needs to be made once.

To pair, starting from the instrument being on, press down the key for a couple of seconds while the Surgic Pro2 device is in pairing mode. "PA" is shown on the display while the pairing is going on. To abort pairing, press the key again.

14.3.2 Sending ISQ value

If the Osseo 100+ is paired with Surgic Pro2, the displayed ISQ and battery status is automatically sent to that device.

15. Cleaning and maintenance



Before use, the parts should be cleaned and disinfected.

15.1 Recommended disinfectants

The following disinfectants are recommended for the instrument, the MultiPeg and the MultiPeg Driver:

- Schülke & Mayr: MikroZid AF Liquid
- Dürr: AF 322
- Metrex: CaviCide

Follow the instructions for use for the disinfectant that is used.



Do not autoclave the instrument

15.2 Autoclave sterilization (MultiPeg & MultiPeg Driver)

Sterilization should be made in a pre-vacuum steam sterilizer (autoclave) according to ISO 17665-1.

Clean the products and put them in an FDA-cleared (USA) autoclave bag before sterilization.

The following sterilization process shall be used:

- At least 3 minutes at 134 (-1/+4)°C or 273(-1.6/+7.4)°F

Follow the instruction for the autoclave that is used.



Do not clean the MultiPeg by ultrasound. Could cause damage



The instrument must be used with a cover in all uses. (Only US)

The instrument must be cleaned with a disinfectant between patients.

15.3 Cleaning

- **Instrument**

The instrument is cleaned with a wet cloth. Any of the above disinfectants can be used. For use in environments requiring sterility, the instrument should be covered with a sterile cover. Recommended covers (US) are Probe Covers, for instance #DYNJE5900 by Medline or part #10040 by Medical Resources.

- **MulTipeg Driver and MulTipegs**

The MulTipeg Driver and the MulTipeg should be cleaned with water with or without detergent, using a light brush. Any of the above disinfectant fluids can be used. For use in environments requiring sterility, the MulTipeg Driver and MulTipeg should be autoclaved before use according to the above instructions.

16. Lifetime

The lifetime of the instrument is determined by the number of charging cycles. The internal batteries can be fully charged more than 500 times before they need to be replaced. The instrument should not be left uncharged for more than 1 year.

The MulTipeg Driver is guaranteed for at least 100 autoclave cycles, and a MulTipeg is guaranteed for at least 20 autoclave cycles, before they are degraded in any way.

17. Troubleshooting

17.1 Possible errors

- **Difficult to achieve a measurement:**

In some cases it is more difficult for the instrument to bring the MulTipeg into vibration. If so, try to hold the instrument tip closer to the tip of the peg. Check also that no soft-tissue is touching the peg which might stop its vibration.

- **Noise warning (audible and visible on the display):**

An electric device close to the instrument is causing the warning. Try to remove the source.

- **The instrument suddenly turns off:**

The instrument turns off automatically after 1 minute of inactivity. It also turns off if the battery level is too low, and due to any of the error codes described below.

- **Not all segments are lit up when instrument is started:**

The instrument is damaged and has to be sent for repair.

17.2 Error codes

If malfunctioning, these error codes are shown on the display before it turns off:

E1: Hardware error. Malfunctioning electronics

E2: Noise error. Shown if constant electromagnetic noise is present

E3: Pulse power error. Malfunctioning magnetic pulse generation



Use of accessories other than those specified or provided by the manufacturer of this equipment could result in increased emissions or decreased electromagnetic immunity of this equipment and result in improper operation.

18. Accessories & Spare Parts

Model	MulTipeg Driver	Sterile Cover	Mains adapter	EU plug	UK plug	AU plug	US plug
REF	55003	55105	55093 incl 1.5 m cable Fuhua Electronic Co., Ltd UE05WCP-052080SPC	55094	55095	55096	55097

MulTipeg: Please refer to the updated list from the supplier.

19. Service

In case of a malfunctioning instrument, the device should be sent to the manufacturer or distributor for repair or exchange.

Osseo 100+ is covered by a two-year warranty.

20. EMC Information

The instrument fulfils the requirements according to EN 60601-1-2 regarding emission and immunity. If sensitive electronic equipment is affected by the instrument, try to increase the distance to such equipment. The charger should not be connected during measurements.

Guidance and manufacturer's declaration - Electromagnetic Emissions.		
Osseo 100 is intended for use in the electromagnetic environment specified below.		
Emissions tests	Compliance	Electromagnetic environment - guidance
RF emissions C1SPR11	Group 1	Osseo 100 uses RF energy only for its internal function.
RF emissions C1SPR11	Class B	Osseo100+ Rechargeable battery operated device
Harmonic emissions IEC61000-3-2	Not applicable	
Voltage fluctuations/flicker emissions IEC61000-3-3	Not applicable	

Guidance and manufacturer's declaration - Electromagnetic Immunity Test Levels		
Osseo100+ is intended for use in the electromagnetic environment specified below.		
Immunity test	EMC standard or test method	Professional healthcare facility environment
Electrostatic discharge (ESD)	IEC61000-4-2	± 8kV contact ± 2 kV ± 4 kV ± 8 kV ± 15 kV air
Radiated RF EM fields	IEC61000-4-3	3 V/m 80 MHz - 2.7 GHz 80 % AM at 1 kHz
Proximity fields form RF wireless communications equipment	IEC61000-4-3	30 cm minimum separation distance from radio transmitter
Rated power frequency magnetic fields	IEC61000-4-8	30 A/m 50 Hz or 60 Hz
Electrical fast transient/burst	IEC 61000-4-4	± 2kV 100 kHz repetition frequency
Surges Line-to-line, Surges Line-to-ground	IEC 61000-4-5	± 0.5, ± 1 kV, ± 2 kV
Conducted disturbances induced by RF fields	IEC61000-4-6	3V 0,15 MHz – 80 MHz 6 V in ISM bands between 0,15 MHz and 80 MHz 80 % AM at 1 kHz
Voltage dips, Voltage interruptions and Electrical transient condition along supply lines	IEC 61000-4-11	5% UT, 0.5 cycle At 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315° 0 % UT; 1 cycle And 70 % UT; 25/30 cycles (50/60Hz) Single phase: at 0° 0 % UT; 250/300 cycle (50/60 Hz)

NAKANISHI INC.

www.nsk-inc.com

700 Shimohinata, Kanuma, Tochigi 322-8666, Japan



Integration Diagnostics Sweden AB
Furstenbergsgatan 4, 416 64 Göteborg, Sweden

NSK Europe GmbH

Elly-Beinhorn-Strasse 8, 6750 Escborn, Germany

NSK France SAS

32 rue de Lisbonne, 75008 Paris, France

NSK United Kingdom Ltd

Office 5, Gateway 1000, Arlington Business Park
Whittle Way, Stevenage, SG1 2FP, UK

NSK Dental Spain SA

C/ Módena, 43 El Soho-Európolis
28232 Las Rozas, Madrid, Spain

NSK America Corp

1800 Global Parkway, Hoffman Estates, IL 60192, USA

NSK America Latina Ltda

Rua Blumenau, 735 – Sala 02, América
Joinville, SC 89204-251, Brazil

NSK Oceania Pty Ltd

Unit 12/809-821 Botany Road
Roseberry, NSW 2018, Australia

NSK Asia

9 Tampines Grande, #03-15, 528735, Singapore

NSK Middle East

Room 6EA701, 7th Floor, East Wing No. 6
Dubai Airport Free Zone, PO Box 54316, Dubai, UAE

Specifications are subject to change without notice.